

# Wide-Bandwidth and High-Precision Current Measurement



HIOKI 9709, CT6862, and CT6863 wide-bandwidth and high-precision AC/DC current sensors achieve high accuracy measurement of both small- and large-current signals. They can be used for a wide variety of applications in the fields of electric and hybrid electric vehicles, inverters and other devices, alternative energy sources such as solar power and fuel cells, as well as home electronic appliances and system equipment.

### Ideal for comprehensive evaluation of wide-bandwidth and high-precision inverter motors



# **HIOKI's Current Sensors Support More Accurate Current and Power Measurements**

#### When a current sensor is used:



New Age for Current Measurement with High Precision, Wide Bandwidth, and Excellent Temperature Characteristics

- Achieve a maximum accuracy of  $\pm 0.16\%$ , which is equivalent to power meters using a direct cable connection (when used with the POWER ANALYZER 3390).
- The cable connecting the power source and load can be made thicker and shorter to allow measurement to be performed in nearly the same cable condition as the real operation of a device under test.
- Insert loss is far smaller than with a shunt resistor so the operation of a device under test is not affected.
- A wide range of measurements are possible, including battery charge and discharge measurement and inverter secondary circuit measurement.
- A wide operating temperature range allows the sensor to be used in the same environment conditions as a device under test.
- Installing a power line is simple, secure, and safe.
- Excellent characteristics reduce the effect of the conductor position, external magnetic field, and magnetization
- CAT III 1000V Safe
- Convenient voltage output design

## Specifications

|  | CT6862  | CT6863                          | 9709  |
|--|---|---------------------------------|---|
| Frequncy characteristics                 | DC to 1 MHz   | DC to 500 kHz                   | DC to 100 kHz   |
| Rated current                            | 50 A AC/DC  | 200 A AC/DC                     | 500 A AC/DC   |
| Output voltage                           | 2 V/50 A  | 2 V/200 A                       | 2 V/500 A   |
| Measurable conductor diameter            | 26 mm(1.0   | 02") or less                    | 36 mm(1.42") or less  |
| Compatible power meters                  | 3193(with 9602), 3194 (with 9602), 3390, PW6001   |                                 |   |
| Output resistance                        | 50 Ω  |                                 |   |
| Basic accuracy                           | DC, $16H z \le f \le 400 Hz$<br>±0.05% rdg. ±0.01% f.s. ±0.2 deg<br>Accuracy is given for 0°c to 40°c, 80% rh or less |                                 | DC, 45 Hz $\leq$ f $\leq$ 66 Hz<br>±0.05% rdg. ±0.01% f.s. ±0.2 deg<br>Accuracy is given for 18°c to 28°c, up to 80% rh |
| Period of accuracy guaranteed            | Accuracy guaranteed for 1 year, Post-adjustment accuracy guaranteed for 1 year  |                                 | Accuracy guaranteed for 6 months,<br>Post-adjustment accuracy guaranteed for 6 months                                   |
| Max. rated voltage to earth              | AC/DC 1000 V (50/60 Hz) (measurement category III )   |                                 |   |
| Cord length                              | 3 m (9.84 ft)   |                                 |   |
| Operating temperature and humidity range |   | 80% RH or less                  | 0°c to 50°c, 80% RH or less<br>(with no condensation)   |
| Dimensions, Mass                         | 70W (2.76") × 100H (3.94") × 5  | 53D (2.09") mm, 340 g (12.0 oz) | $160W (6.30") \times 112H (4.41") \times 50D (1.97") mm, 850 g(30.0 oz)$  |
| Supplied Accessories                     | Instruction manual ×1, Mark bands ×6  |                                 |   |





#### Mounting diagram (unit : mm)





| Model : AC/DC CURRENT SENSOR<br>Model No. (Order Code) (Note) |                                |  |  |
|---|--------------------------------|--|--|
|   |                                |  |  |
| CT6862-05   | (50 A AC/DC, 12 pin terminal)  |  |  |
| CT6863  | (200 A AC/DC)                  |  |  |
| CT6863-05   | (200 A AC/DC, 12 pin terminal) |  |  |
| 9709  | (500A AC/DC)                   |  |  |
| 9709-05   | (500 A AC/DC, 12 pin terminal) |  |  |

HIOKI E.E. CORPORATION

#### HEADQUARTERS

81 Koizumi, Ueda, Nagano, 386-1192, Japan TEL +81-268-28-0562 FAX +81-268-28-0568 http://www.hioki.com / E-mail: os-com@hioki.co.jp

#### HIOKI USA CORPORATION

TEL +1-609-409-9109 FAX +1-609-409-9108 http://www.hiokiusa.com / E-mail: hioki@hiokiusa.com

All information correct as of Sept. 29, 2016. All specifications are subject to change without notice

HIOKI (Shanghai) SALES & TRADING CO., LTD. TEL +86-21-63910090 FAX +86-21-63910360

http://www.hioki.cn / E-mail: info@hioki.com.cn

HIOKI SINGAPORE PTE. LTD.

TEL +65-6634-7677 FAX +65-6634-7477 E-mail: info-sg@hioki.com.sg

HIOKI KOREA CO., LTD.

TEL +82-2-2183-8847 FAX +82-2-2183-3360 E-mail: info-kr@hioki.co.jp

DISTRIBUTED BY